

SOLUTION CHANGE

GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306

Date: 06/06/2021

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:01
INTERNAL STANDARD	VERIFIED	03:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:02
BLANK TEST	0.00	03:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:03
BLANK TEST	0.00	03:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:04
BLANK TEST	0.00	03:05

