

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 06/20/2021

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:10
INTERNAL STANDARD	VERIFIED	05:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:11
BLANK TEST	0.00	05:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	05:12
BLANK TEST	0.00	05:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	05:13
BLANK TEST	0.00	05:14

