

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 06/19/2021

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:10
INTERNAL STANDARD	VERIFIED	07:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:10
BLANK TEST	0.00	07:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:12
BLANK TEST	0.00	07:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:13
BLANK TEST	0.00	07:14

