

SUPERVISOR TEST

 CAMBRIDGE VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 126706

Date: 06/21/2021

OPERATOR NAME:
 ROBERT M DANKO
 Permit #: 39105
 Operator Agency/Dept: CAMBRIDGE VILL. PD
 Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:10
INTERNAL STANDARD	VERIFIED	07:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:10
BLANK TEST	0.00	07:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:11
BLANK TEST	0.00	07:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:13
BLANK TEST	0.00	07:13

