

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 06/27/2021

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:51
INTERNAL STANDARD	VERIFIED	03:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:51
BLANK TEST	0.00	03:52
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:53
BLANK TEST	0.00	03:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:54
BLANK TEST	0.00	03:54

