

SOLUTION CHANGE

-----  
GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306  
-----

Date: 07/03/2021

OPERATOR NAME:  
RYAN S PEDONE  
Permit #: 35878  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:53
INTERNAL STANDARD	VERIFIED	21:53
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	21:54
BLANK TEST	0.00	21:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:55
BLANK TEST	0.00	21:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	21:56
BLANK TEST	0.00	21:57

