

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 06/26/2021

OPERATOR NAME:
KYLE L KOLAR
Permit #: 44063
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:19
INTERNAL STANDARD	VERIFIED	21:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:20
BLANK TEST	0.00	21:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:21
BLANK TEST	0.00	21:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	21:23
BLANK TEST	0.00	21:24

2 #691
06/26/2021

