

SUPERVISOR TEST

 WASHINGTON CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100284

Date: 06/28/2021

OPERATOR NAME:
 ROBERT J SULLIVAN
 Permit #: 34106
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:09
INTERNAL STANDARD	VERIFIED	17:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:10
BLANK TEST	0.00	17:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:11
BLANK TEST	0.00	17:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:12
BLANK TEST	0.00	17:13

