

SUPERVISOR TEST

HUDSON FALLS  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 102408

Date: 07/07/2021

OPERATOR NAME:  
 JOSEPH F MCKEEVER  
 Permit #: 49488  
 Operator Agency/Dept: HUDSON FALLS VIL. PD  
 Reference Standard #: 53443

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:05
INTERNAL STANDARD	VERIFIED	22:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:05
BLANK TEST	0.00	22:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	22:06
BLANK TEST	0.00	22:07
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:08
BLANK TEST	0.00	22:08

