

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dnt: 113806

Date: 07/10/2021

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:21
INTERNAL STANDARD	VERIFIED	08:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:21
BLANK TEST	0.00	08:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:23
BLANK TEST	0.00	08:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:24
BLANK TEST	0.00	08:25

