

SUPERVISOR TEST

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 06/26/2021

OPERATOR NAME:  
BRYAN J GRECO  
Permit #: 44077  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:23
INTERNAL STANDARD	VERIFIED	18:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:24
BLANK TEST	0.00	18:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:25
BLANK TEST	0.00	18:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	18:26
BLANK TEST	0.00	18:27

