

SUPERVISOR TEST

HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 07/21/2021

OPERATOR NAME:
 JOSEPH F MCKEEVER
 Permit #: 49488
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 53443

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:45
INTERNAL STANDARD	VERIFIED	22:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:46
BLANK TEST	0.00	22:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	22:47
BLANK TEST	0.00	22:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	22:48
BLANK TEST	0.00	22:49

