

SUPERVISOR TEST

HUDSON FALLS  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 102408

Date: 06/20/2021

OPERATOR NAME:  
 JOSEPH F MCKEEVER  
 Permit #: 49488  
 Operator Agency/Dept: HUDSON FALLS VIL. PD  
 Reference Standard #: 53443

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:02
INTERNAL STANDARD	VERIFIED	03:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	03:03
BLANK TEST	0.00	03:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:04
BLANK TEST	0.00	03:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	03:05
BLANK TEST	0.00	03:06

