

SUPERVISOR TEST

GRANVILLE VILLAGE  
 POLICE DEPT  
 DATAMASTER dmt: 113306

Date: 07/25/2021

OPERATOR NAME:  
 RYAN S PEDONE  
 Permit #: 35878  
 Operator Agency/Dept: GRANVILLE VILLAGE PD  
 Reference Standard #: 21100

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 04:07 |
| INTERNAL STANDARD       | VERIFIED | 04:07 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.10     | 04:08 |
| BLANK TEST              | 0.00     | 04:09 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.10     | 04:09 |
| BLANK TEST              | 0.00     | 04:10 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.10     | 04:10 |
| BLANK TEST              | 0.00     | 04:11 |

