

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 07/15/2021

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:21
INTERNAL STANDARD	VERIFIED	12:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:22
BLANK TEST	0.00	12:23
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:23
BLANK TEST	0.00	12:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:24
BLANK TEST	0.00	12:25

