

SUPERVISOR TEST

CAMBRIDGE VILLAGE
POLICE DEPT
DATAMASTER dmt: 126706

Date: 07/26/2021

OPERATOR NAME:
ROBERT M DANKO
Permit #: 39105
Operator Agency/Dept: CAMBRIDGE VILL. PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:55
INTERNAL STANDARD	VERIFIED	18:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:55
BLANK TEST	0.00	18:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:56
BLANK TEST	0.00	18:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	18:57
BLANK TEST	0.00	18:58

