

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 07/31/2021

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:28
INTERNAL STANDARD	VERIFIED	22:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:28
BLANK TEST	0.00	22:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:30
BLANK TEST	0.00	22:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:31
BLANK TEST	0.00	22:32

