

SUPERVISOR TEST

HUDSON FALLS  
 VILLAGE POLICE DEPT.  
 DATAMASTER dmt: 102408

Date: 07/28/2021

OPERATOR NAME:  
 JOSEPH F MCKEEVER  
 Permit #: 49488  
 Operator Agency/Dept: HUDSON FALLS VIL. PD  
 Reference Standard #: 53443

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:58
INTERNAL STANDARD	VERIFIED	08:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:59
BLANK TEST	0.00	08:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:00
BLANK TEST	0.00	09:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:01
BLANK TEST	0.00	09:02

