

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 08/05/2021

OPERATOR NAME:
JEFFREY E PETERMAN
Permit #: 49748
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:48
INTERNAL STANDARD	VERIFIED	14:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:49
BLANK TEST	0.00	14:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:50
BLANK TEST	0.00	14:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:51
BLANK TEST	0.00	14:52

