

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 08/22/2021

OPERATOR NAME:
DALE W QUESNEL
Permit #: 37510
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:14
INTERNAL STANDARD	VERIFIED	19:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	19:14
BLANK TEST	0.00	19:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	19:16
BLANK TEST	0.00	19:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:17
BLANK TEST	0.00	19:18

