

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 08/09/2021

OPERATOR NAME:
MATTHEW K JACKSON
Permit #: 40078
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:37
INTERNAL STANDARD	VERIFIED	23:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:38
BLANK TEST	0.00	23:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:39
BLANK TEST	0.00	23:40
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:40
BLANK TEST	0.00	23:41

