

SUPERVISOR TEST

FORT EDWARD
VILLAGE POLICE DEPT.
DATAMASTER dmt: 115708

Date: 08/09/2021

OPERATOR NAME:
SGT DEAN E WATKINS
Permit #: 39523
Operator Agency/Dept: FORT EDWARD VILL. PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:25
INTERNAL STANDARD	VERIFIED	23:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	23:26
BLANK TEST	0.00	23:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:27
BLANK TEST	0.00	23:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:28
BLANK TEST	0.00	23:29

