

SUPERVISOR TEST

 FORT EDWARD
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 115708

Date: 08/16/2021

OPERATOR NAME:
 SGT DEAN E WATKINS
 Permit #: 39523
 Operator Agency/Dept: FORT EDWARD VILL. PD
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:32
INTERNAL STANDARD	VERIFIED	23:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:32
BLANK TEST	0.00	23:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:33
BLANK TEST	0.00	23:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:35
BLANK TEST	0.00	23:36

