

SUPERVISOR TEST

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 08/23/2021

OPERATOR NAME:  
SGT DEAN E WATKINS  
Permit #: 39523  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:56
INTERNAL STANDARD	VERIFIED	22:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:56
BLANK TEST	0.00	22:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:58
BLANK TEST	0.00	22:59
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	22:59
BLANK TEST	0.00	23:00

