

SUPERVISOR TEST

WASHINGTON COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106406

Date: 08/24/2021

OPERATOR NAME:
 CORY L HURLBURT
 Permit #: 37017
 Operator Agency/Dept: WASHINGTON CO. SO
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:37
INTERNAL STANDARD	VERIFIED	06:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:38
BLANK TEST	0.00	06:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	06:39
BLANK TEST	0.00	06:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:40
BLANK TEST	0.00	06:41

Hurlburt # 257

