

SOLUTION CHANGE


WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 09/01/2021

OPERATOR NAME:
CHRISTOPHER J MURRAY
Permit #: 46226
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:16
INTERNAL STANDARD	VERIFIED	17:16
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:17
BLANK TEST	0.00	17:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:18
BLANK TEST	0.00	17:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:19
BLANK TEST	0.00	17:20



#821

