

SUPERVISOR TEST

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 08/30/2021

OPERATOR NAME:
MICHAEL O WEBER
Permit #: 46067
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:41
INTERNAL STANDARD	VERIFIED	07:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:42
BLANK TEST	0.00	07:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:43
BLANK TEST	0.00	07:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:44
BLANK TEST	0.00	07:45

