

SOLUTION CHANGE

WASHINGTON CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100284

Date: 09/01/2021

OPERATOR NAME:  
ROBERT J SULLIVAN  
Permit #: 34106  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:03
INTERNAL STANDARD	VERIFIED	21:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:04
BLANK TEST	0.00	21:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:06
BLANK TEST	0.00	21:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	21:07
BLANK TEST	0.00	21:08

