

SUPERVISOR TEST

-----  
 CAMBRIDGE VILLAGE  
 POLICE DEPT  
 DATAMASTER dmt: 126706  
 -----

Date: 09/14/2021

OPERATOR NAME:  
 ROBERT M DANKO  
 Permit #: 39105  
 Operator Agency/Dept: CAMBRIDGE VILL. PD  
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:10
INTERNAL STANDARD	VERIFIED	07:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:10
BLANK TEST	0.00	07:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:12
BLANK TEST	0.00	07:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:13
BLANK TEST	0.00	07:13

