

SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 09/04/2021

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:42
INTERNAL STANDARD	VERIFIED	20:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:42
BLANK TEST	0.00	20:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:44
BLANK TEST	0.00	20:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	20:45
BLANK TEST	0.00	20:45

