

SUPERVISOR TEST

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 09/11/2021

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35787
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:55
INTERNAL STANDARD	VERIFIED	03:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:55
BLANK TEST	0.00	03:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:57
BLANK TEST	0.00	03:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:58
BLANK TEST	0.00	03:59

