

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 09/06/2021

OPERATOR NAME:
WILLIAM M GREENE
Permit #: 48760
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:51
INTERNAL STANDARD	VERIFIED	14:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:51
BLANK TEST	0.00	14:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:52
BLANK TEST	0.00	14:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:54
BLANK TEST	0.00	14:55

