

SUPERVISOR TEST

HUDSON FALLS
VILLAGE POLICE DEPT.
DATAMASTER dmt: 102408

Date: 09/07/2021

OPERATOR NAME:
DAVID M COSTELLO
Permit #: 40891
Operator Agency/Dept: HUDSON FALLS VIL. PD
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:59
INTERNAL STANDARD	VERIFIED	04:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	04:59
BLANK TEST	0.00	05:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	05:00
BLANK TEST	0.00	05:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	05:01
BLANK TEST	0.00	05:02

