

SOLUTION CHANGE

 HUDSON FALLS
 VILLAGE POLICE DEPT.
 DATAMASTER dmt: 102408

Date: 09/13/2021

OPERATOR NAME:
 DAVID M COSTELLO
 Permit #: 40891
 Operator Agency/Dept: HUDSON FALLS VIL. PD
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:47
INTERNAL STANDARD	VERIFIED	19:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:48
BLANK TEST	0.00	19:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:49
BLANK TEST	0.00	19:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	19:50
BLANK TEST	0.00	19:51

