

SOLUTION CHANGE

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 08/10/2021

OPERATOR NAME:  
DAVID F BUXTON  
Permit #: 33398  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:38
INTERNAL STANDARD	VERIFIED	18:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:39
BLANK TEST	0.00	18:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:40
BLANK TEST	0.00	18:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	18:41
BLANK TEST	0.00	18:42

