

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 08/23/2021

OPERATOR NAME:
BRYAN J GRECO
Permit #: 44077
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:00
INTERNAL STANDARD	VERIFIED	22:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:01
BLANK TEST	0.00	22:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:02
BLANK TEST	0.00	22:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	22:03
BLANK TEST	0.00	22:04

