

SOLUTION CHANGE

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 09/04/2021

OPERATOR NAME:  
ALEXANDER T GILMORE  
Permit #: 47649  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:09
INTERNAL STANDARD	VERIFIED	06:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:10
BLANK TEST	0.00	06:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:11
BLANK TEST	0.00	06:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	06:12
BLANK TEST	0.00	06:13

