

SUPERVISOR TEST

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 09/20/2021

OPERATOR NAME:  
NICHOLAS D HOFFMAN  
Permit #: 49232  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:09
INTERNAL STANDARD	VERIFIED	16:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	16:10
BLANK TEST	0.00	16:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	16:11
BLANK TEST	0.00	16:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	16:12
BLANK TEST	0.00	16:13

