

SUPERVISOR TEST

-----  
 WASHINGTON CO.  
 SHERIFF'S OFFICE  
 DATAMASTER dmt: 100284  
 -----

Date: 09/20/2021

OPERATOR NAME:  
 ERIC C WRIGHT  
 Permit #: 49047  
 Operator Agency/Dept: WASHINGTON CO. SO  
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:37
INTERNAL STANDARD	VERIFIED	17:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.08	17:37
BLANK TEST	0.00	17:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.08	17:38
BLANK TEST	0.00	17:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.08	17:39
BLANK TEST	0.00	17:40

