

SUPERVISOR TEST

-----  
FORT EDWARD  
VILLAGE POLICE DEPT.  
DATAMASTER dmt: 115708  
-----

Date: 09/20/2021

OPERATOR NAME:  
SGT DEAN E WATKINS  
Permit #: 39523  
Operator Agency/Dept: FORT EDWARD VILL. PD  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:08
INTERNAL STANDARD	VERIFIED	23:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:08
BLANK TEST	0.00	23:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:10
BLANK TEST	0.00	23:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	23:11
BLANK TEST	0.00	23:12

