

SOLUTION CHANGE

WASHINGTON CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 100284

Date: 09/27/2021

OPERATOR NAME:
JOSHUA T WHITNEY
Permit #: 40382
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:08
INTERNAL STANDARD	VERIFIED	17:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:09
BLANK TEST	0.00	17:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:10
BLANK TEST	0.00	17:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:11
BLANK TEST	0.00	17:12

