

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 09/25/2021

OPERATOR NAME:  
ALEXANDER T GILMORE  
Permit #: 47649  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:48
INTERNAL STANDARD	VERIFIED	21:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:49
BLANK TEST	0.00	21:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:50
BLANK TEST	0.00	21:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	21:51
BLANK TEST	0.00	21:52

