

SUPERVISOR TEST

WHITEHALL VILLAGE
POLICE DEPT.
DATAMASTER dmt: 113806

Date: 09/18/2021

OPERATOR NAME:
ALEXANDER T GILMORE
Permit #: 47649
Operator Agency/Dept: WHITEHALL VILL. P.D.
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:56
INTERNAL STANDARD	VERIFIED	17:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:56
BLANK TEST	0.00	17:57
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:57
BLANK TEST	0.00	17:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:58
BLANK TEST	0.00	17:59

