

SOLUTION CHANGE

 CAMBRIDGE VILLAGE
 POLICE DEPT
 DATAMASTER dmt: 126706

Date: 10/01/2021

OPERATOR NAME:
 ROBERT M DANKO
 Permit #: 39105
 Operator Agency/Dept: CAMBRIDGE VILL. PD
 Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:01
INTERNAL STANDARD	VERIFIED	14:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:02
BLANK TEST	0.00	14:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:03
BLANK TEST	0.00	14:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:04
BLANK TEST	0.00	14:05

