

SOLUTION CHANGE

GRANVILLE VILLAGE
POLICE DEPT
DATAMASTER dmt: 113306

Date: 10/03/2021

OPERATOR NAME:
RYAN S PEDONE
Permit #: 35878
Operator Agency/Dept: GRANVILLE VILLAGE PD
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:01
INTERNAL STANDARD	VERIFIED	06:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:01
BLANK TEST	0.00	06:02
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:02
BLANK TEST	0.00	06:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	06:04
BLANK TEST	0.00	06:04

