

SOLUTION CHANGE

WASHINGTON COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106406

Date: 10/01/2021

OPERATOR NAME:  
WILLIAM M GREENE  
Permit #: 487602  
Operator Agency/Dept: WASHINGTON CO. SO  
Reference Standard #: 21280

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 21:19 |
| INTERNAL STANDARD       | VERIFIED | 21:19 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.10     | 21:19 |
| BLANK TEST              | 0.00     | 21:20 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.10     | 21:21 |
| BLANK TEST              | 0.00     | 21:22 |
| REFERENCE STANDARD TEMP | 33.99c   |       |
| REFERENCE STANDARD      | 0.10     | 21:22 |
| BLANK TEST              | 0.00     | 21:23 |

*WJG #647*

