

SUPERVISOR TEST

GRANVILLE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 113306

Date: 10/19/2021

OPERATOR NAME:  
RONALD D TAYLOR III  
Permit #: 49003  
Operator Agency/Dept: GRANVILLE VILLAGE PD  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:23
INTERNAL STANDARD	VERIFIED	03:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:23
BLANK TEST	0.00	03:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:24
BLANK TEST	0.00	03:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:25
BLANK TEST	0.00	03:26

