

SUPERVISOR TEST

WASHINGTON COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 106406

Date: 10/11/2021

OPERATOR NAME:
ROBERT J SULLIVAN
Permit #: 34106
Operator Agency/Dept: WASHINGTON CO. SO
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:50
INTERNAL STANDARD	VERIFIED	17:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	17:50
BLANK TEST	0.00	17:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:52
BLANK TEST	0.00	17:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	17:53
BLANK TEST	0.00	17:54

