

SUPERVISOR TEST

WHITEHALL VILLAGE  
POLICE DEPT.  
DATAMASTER dmt: 113806

Date: 10/17/2021

OPERATOR NAME:  
ALEXANDER T GILMORE  
Permit #: 47649  
Operator Agency/Dept: WHITEHALL VILL. P.D.  
Reference Standard #: 21280

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:11
INTERNAL STANDARD	VERIFIED	00:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:12
BLANK TEST	0.00	00:13
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:13
BLANK TEST	0.00	00:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	00:14
BLANK TEST	0.00	00:15

