

SUPERVISOR TEST

-----  
CAMBRIDGE VILLAGE  
POLICE DEPT  
DATAMASTER dmt: 126706  
-----

Date: 10/11/2021

OPERATOR NAME:  
ROBERT M DANKO  
Permit #: 39105  
Operator Agency/Dept: CAMBRIDGE VILL. PD  
Reference Standard #: 21170

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:54
INTERNAL STANDARD	VERIFIED	13:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:55
BLANK TEST	0.00	13:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:56
BLANK TEST	0.00	13:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:57
BLANK TEST	0.00	13:58

